


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P-PS*
S.N.: *1152*
Date: *JAN 20 83*



- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 25 1000 Hz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV.
1.0 nSEC/DIV.
4.0 MHz.

110/220 V,
50-60 Hz.

[Signature]