

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


AV6-2-P

S.N.:

1154

Date:

JAN 20 83



a) Output signal amplitude:

0 to 100 V

b) Pulse width:

c) Jitter, stability:

d) Prime power:

+15 ± 1.0 V

200 mA (max)

100 VOLTS/DIV

2.0 NSEC/DIV

20 KHZ