


PULSE GENERATOR

PERFORMANCE CHECK

Model: AUP-AU-1-N-PS

S.N.: 1158

Date: DEC 30 82



- a) Output signal amplitude:
0 TO -10V.
- b) Pulse width:
0 TO 4 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
400 KHz.

120/240V -
50-60 Hz.

UNIT SHIPPED
WITH VOLTAGE
SELECTION CARD
IN THE 240V
POSITION.