


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PTV-1A-1-P*

S.N.: *1162*

Date: *JAN 20 83*



*100 nSEC/DIV.
50 VOLTS/DIV
R_L = 50 Ω.
1.5 KHz.*

- a) Output signal amplitude:
0 TO +200V
- b) Pulse width:
50 nSEC TO 1.0 μsec
- c) Rise time: (*SEE FIG 1*)
≤ 10 nSEC
- d) Fall time:
≤ 10 nSEC
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability: (*SEE FIG 1*)
- g) Prime power:
OK

*+24 ± 1.0 V
400 mA (max)*