


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-1-PW

S.N.: 1164

Date: FEB 17 83



- a) Output signal amplitude:  
0 to +200 V
- b) Pulse width:  
100 nsec to 5  $\mu$ sec
- c) Rise time:  
 $\leq 10$  nsec
- d) Fall time:  
 $\leq 20$  nsec
- e) PRF:  
0 to 1 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

110/220V,  
50-60 Hz.

50 VOLTS/DIV  
200 nsec/DIV  
1 kHz