


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVP-AV-15-N-ECL-C-05*

S.N.: *1165*

Date: *MAR 10 83*



*5 VOLTS/DIV
1.0 NSEC/DIV
0.1 MHz.*



*1.0 VOLTS/DIV
100 PSEC/DIV
0.1 MHz*

- a) Output signal amplitude:
0 TO -5 V.
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

*20/240V
50-60 Hz.*

TO 08799 ECL OUTPUT

