


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-L1

S.N.: 1166

Date: FEB 9 83



a) Output signal amplitude:

$\approx +40V$

b) Pulse width:

1 TO 100 NSEC

c) Rise time:

$\leq 300$  PSEC

d) Fall time:

$\leq 2$  NSEC

e) PRF:

SEE FIG 1 & 2

f) Jitter, stability:

OK.

g) Prime power:

+15V  $\pm$  1.0V

70 mA MAX.

1 NSEC/DIV  
32 VOLTS/DIV  
40 KHZ.