


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI-4*
S.N.: *1168*
Date: *FEB 9 83*



- a) Output signal amplitude:
> 40V
- b) Pulse width:
1 TO 100 NSEC
- c) Rise time:
≤ 300 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE FILE 142
- f) Jitter, stability:
OK
- g) Prime power:

+15V ± 1.0V

70 mA (MAX)

*1 NSEC/DIV
32 VOLTS/DIV
40 KHz*

[Signature]