


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-L1*  
S.N.: *1169*  
Date: *FEB 9 83*



*32 VOLTS/DIV*  
*1 NS/2 DIV*  
*40 kHz*

- a) Output signal amplitude:  
*> 140 V*
- b) Pulse width:  
*1 TO 100 NS/2*
- c) Rise time:  
*≤ 300 PSEZ*
- d) Fall time:  
*≤ 2 NS/2*
- e) PRF:  
*SEE FILE 182*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 V ± 1.0 V*  
*70 mA (max)*

*[Signature]*