


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-41*  
S.N.: *1171*  
Date: *FEB 9 83*



- a) Output signal amplitude:  
*740 VOLTS*
- b) Pulse width:  
*1 TO 100 NSEC*
- c) Rise time:  
*≤ 300 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*SEE FIG 1 & 2*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*32 VOLTS / DIV*  
*1 NSEC / DIV*  
*40 kHz*

*+15V ± 1.0V*  
*70 mA (MAX)*