


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV0-5C-PS-015

S.N.: 1173

Date: MAR 28 83



≈ 10 AMPS/DIV
10 NSEC/DIV
10 KHZ.
($R_L = 3 \Omega$)

- a) Output signal amplitude: *
- b) Pulse width: (TO 3 Ω)
- c) Rise time:
- d) Fall time:
- e) PRF:
- f) Jitter, stability:
- g) Prime power:

0 TO +10 AMPS
25 TO 120 NSEC
≤ 5 NSEC
≤ 5 NSEC
0 TO 10 KHZ
OK
110/220 V
50-60 HZ

* I_{out} CONTROLLED BY DC CONTROL VOLTAGE, 0 TO +10 V. SEE FILE 1.