

PULSE GENERATOR

PERFORMANCE CHECK

Model:


*AVR-A-1-PS-CA-EN-P*

S.N.:

*1174*

Date:

*FEB 15 83*



- a) Output signal amplitude:  
*0 TO 1200V*
- b) Pulse width:  
*70 NSEC TO 1.0 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 20 NSEC.*
- e) PRF:  
*0 TO 5 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 VOLTS/DIV*

*100 NSEC/DIV*

*2 KHZ.*

*R<sub>L</sub> = 50 Ω*

*110/220V*  
*50-60 HZ.*

