


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVUM-2-N-PS

S.N.: 1178

Date: FEB 24 63



- a) Output signal amplitude:
 $\approx -15V$
- b) Pulse width:
 0 TO 2 nsec
- c) Rise time:
 ≤ 100 psec
- d) Fall time:
 ≤ 135 nsec
- e) PRF:
 0 TO 25 MHz
- f) Jitter, stability:
 OK
- g) Prime power: *

10 VOLTS/DIV
1.0 nsec/DIV
4.0 MHz

120/240V
50-60 Hz
UNIT SHIPPED
WITH VOLTAGE
SELECTOR CHANGED
IN THE 240V
POSITION.