


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-101

S.N.: 1179

Date: JAN 19 83



a) Output signal amplitude:

$\geq +30V$

b) Pulse width:

$\leq 1 \text{ nSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+135 \pm 5 \text{ VOLTS}$

$5 \text{ mA (MAX)}$

16 VOLTS/DIV

1.0 nSEC/DIV

400 kHz.