


PULSE GENERATOR

PERFORMANCE CHECK

Model: 17VK-AV-N-TTL-05-PS

S.N.: 1132

Date: Feb 28 83



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 2 NSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

2 NSEC/DIV
10 VOLTS/DIV
20 KHZ

120/240V
50-60 HZ