


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-1-N-TT2-05-PS

S.N.: 1183

Date: Feb 28 83



a) Output signal amplitude:

0 TO - 10V

b) Pulse width:

0 TO 4 NSER

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240V

50-60 Hz.

10 VOLTS/DIV

100 NSER/DIV

400 KHz.