


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANI-V-A-N

S.N.: 1184

Date: Feb 28 83



5 nsec/div
32 volts/div
20 kHz

- a) Output signal amplitude:
0 to 50V
- b) Pulse width:
0 to 100 nsec
- c) Rise time:
≤ 500 psec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15V ± 1.0V
200 mA