


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-N-N-A*

S.N.: *1197*

Date: *DEC 28 83*



- a) Output signal amplitude:
0 TO -50V
- b) Pulse width:
0 TO 100 nSEC
- c) Rise time:
≤ 0.5 nSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*5 nSEC/DIV
32 VOLTS/DIV
20 KHz.*

+15 V ± 1.0V

220 mA.

[Signature]