


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-AV-P*

S.N.: *1201*

Date: *MAR 10 83*



*10 VOLTS/DIV.*

*2.0 NSEC/DIV.*

*40 KHZ.*

- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*0 TO 50 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 50 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15V ± 1.0V.*

*225 mA (MAX)*