


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-1-N

S.N.: 1202

Date: APRIL 7 83



- a) Output signal amplitude:
0 TO -200 V TO 50 Ω
- b) Pulse width:
100 NSEC TO 5 μ sec.
- c) Rise time:
 \leq 10 NSEC
- d) Fall time:
 \leq 20 NSEC
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

50 VOLTS/DIV.

100 NSEC/DIV

1.0 KHz

$R_L = 50 \Omega$.

120 / 240 V.
50 - 60 Hz.

NY