


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PK-A-7*

S.N.: *1267*

Date: *MAR 25 83*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
4 TO 100 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC.
- e) PRF:
SEE FIG 1+2
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V

200 mA (MAX)

2 NSEC/DIV
10 VOLTS/DIV
40 KHZ