


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-A-N*

S.N.: *1209*

Date: *FEB 7 83*



a) Output signal amplitude:

*0 to -32 V*

b) Pulse width:

*≤ 1 nsec*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 V*

*150 mA (MAX)*

*16 VOLTS/DIV*

*100 nsec/DIV*

*400 kHz*