


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANP-MV-1-D-PS-05-TTR

S.N.: 1215

Date: JUN 28 83



- a) Output signal amplitude:
0 TO +10 V
- b) Pulse width:
0 TO 4 NS/2
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK*
- g) Prime power:
120/240 V
50-60 Hz

10 VOLTS/DIV.
1.0 NS/2/DIV
400 kHz

* NOTE: UNIT SHIPPED
WITH VOLTAGE SELECTOR
CARD IN THE 240 V
POSITION