


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-N*
S.N.: *1217*
Date: *MAY 9 83*



32 VOLTS/DIV.
2.0 NSEC/DIV.

- a) Output signal amplitude:
≥ -40V
- b) Pulse width: *SEE FIG 1+2.*
1 TO 100 NSEC.
- c) Rise time:
≤ 300 PSEC
- d) Fall time:
≤ 2 NSEC.
- e) PRF:
SEE FIG 1+2
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT
150 mA (MAX)
[Signature]