


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-A-P*
S.N.: *1224*
Date: *FEB 13 83*



- a) Output signal amplitude:
0 to +50V
- b) Pulse width:
0 to 100 nsec
- c) Rise time:
≤ 500 psec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

~32 VOLTS/DIV
2.0 nsec/DIV
20 kHz

+15 ± 1.0V
200 mA
(MAX)
[Signature]