


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-A-P*  
S.N.: *1226*  
Date: *MAR 25 83*



- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*4 TO 100 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC.*
- e) PRF:  
*SEE FIG 1 & 2.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 V*  
*200 mA (MAX)*

*2 NSEC / DIV*  
*10 VOLTS / DIV*  
*40 KHZ.*