


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT

S.N.: 1228

Date: FEB 7 83



a) Output signal amplitude:

$\geq +30 \text{ V}$

b) Pulse width:

$\leq 1 \text{ NSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ V}$

150 mA (MAX)

32 VOLTS/DIV.  
1.0 NSEC/DIV.

400 KHz