


PULSE GENERATOR
PERFORMANCE CHECK

852078

Model: *PI-4V2*

S.N.: *1245*

Date: *MAR 22 83*



- a) Output signal amplitude:
≈ 90 V
- b) Pulse width:
2.3 ± 0.3 nsec
- c) Rise time:
≈ 580 psec
- d) Fall time:
≈ 1 nsec
- e) PRF:
0 TO 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:

50 VOLTS/DIV
1.0 nsec/DIV
40 kHz

+15 ± 1.0 VOLT
200 mA.