


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2-N-PS*

S.N.: *1249*

Date: *MAR 8 83*



a) Output signal amplitude:

*0 to -350 V*

b) Pulse width:

*0 to 100 nsec*

c) Rise time:

*≤ 1.5 nsec*

d) Fall time:

*≤ 2 nsec*

e) PRF:

*0 to 5 kHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240 V*

*50 - 60 Hz*

*2 320 VOLTS/DIV*

*10 nsec/DIV*

*2 kHz*