

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


A20-3B-P

S.N.:

1253

Date:

FEB 13 83



a) Output signal amplitude:

+10 TO +30 AMPS

b) Pulse width:

≈ 1.5 NSEC.

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V

125 mA (MAX)

*[Signature]*

≈ 20 AMP/DIV

2.0 NSEC