


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-P-PS*

S.N.: *1261*

Date: *FEB 24 83*



- a) Output signal amplitude:  
*0 TO +15 V*
- b) Pulse width:  
*0 TO 2 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV  
1.0 NSEC/DIV  
4.0 MHz.*

*120/240 V  
50 - 60 Hz.*