


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-AU-P-T12.*

S.N.: *1263*

Date: *MAR 30 83*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

2.0 NSEC/DIV
10 VOLTS/DIV
40 KHz

+15 ± 1.0V
225 mA (max)
[Signature]