


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-AU-D-TTL*

S.N.: *1264*

Date: *MAR 30 63*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 58 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V
210 mA (max)
[Signature]

2.0 nSEC/DIV
10 VOLTS/DIV
40 kHz