


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PVI-V-14VI-C-PN*

S.N.: *1268*

Date: *MAY 31 83*



*1.0 NSEC/DIV.  
16 VOLTS/DIV.*



*5 NSEC/DIV.  
32 VOLTS/DIV.*

*(110 KHz)*

- a) Output signal amplitude:  
*0 TO ± 50 VOLTS.*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 250 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*110/220V*

*50-60 Hz.*

*[Signature]*