


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVN-5-1-N-PS

S.N.: 1270

Date: APRIL 28 83



- a) Output signal amplitude:
0 TO -5V
- b) Pulse width:
0.2 TO 1.0 NSEC
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 135 PSEC
- e) PRF:
50 - 260 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV.
1.0 NSEC/DIV.
100 MHz.

110/220 V
50 - 60 MHz