

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVK-AU-P-TTR-PS

S.N.:

1271

Date:

MAR 30 63



a) Output signal amplitude:

0 TO +15V

b) Pulse width:

0 TO 50 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 50 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50 - 60 Hz

2.0 NSEC/DIV
10 VOLTS/DIV
40 KHz