


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-HV3-C-P*  
S.N.: *1276*  
Date: *MAY 31 83*



- a) Output signal amplitude:  
*0 TO +40V*
- b) Pulse width:  
*0 TO 2 NSEC*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 250 PSEC*
- e) PRF:  
*0 TO 1 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*16 VOLTS/DIV.*  
*1.0 NSEC/DIV.*  
*100 KHZ.*

*110/220V*  
*50-60 Hz.*

*[Signature]*