


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-AV-N-TTL*

S.N.: *1277*

Date: *APRIL 25 83*



*10 VOLTS/DIV
2.0 NSEC/DIV
40 KHZ.*

- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V

225 mA