


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2-P*

S.N.: *1278*

Date: *APRIL 25 1983*



*≈ 3.5 AMPS / DIV  
5 NSEC / DIV  
10 KHZ.*

- a) Output signal amplitude:  
*0 TO +5 AMPS*
- b) Pulse width:  
*0 TO 50 NSEC*
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO ~~20~~ 20 KHZ*
- f) Jitter, stability:
- g) Prime power:

*+15 ± 1.0 VOLT*

*230 mA.*

