


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-1-PW-N*

S.N.: *1281*

Date: *MAY 16 83*



- a) Output signal amplitude:
0 TO -200V
- b) Pulse width:
0.1 TO 5 USEC.
- c) Rise time:
≤ 20 NSEC
- d) Fall time:
≤ 20 NSEC
- e) PRF:
0 TO 1 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

~~100 V~~
50 VOLTS/DIV.
1.0 USEC/DIV
400 HZ.

110/220V
50-60 HZ.