


IMPULSE GENERATOR

PERFORMANCE CHECK


Model: *AVG-1-CU-2*

S.N.: *1282*

Date: *JUNE 8 83*



Ⓐ *AVG* OUTPUT TO *50 Ω*
100 VOLTS/DIV
2.0 NSEC/DIV
2 KHZ.



Ⓑ *AS* Ⓐ BUT *AVX-M3-CU*
OUTPUT TO *12.5 Ω*
100 VOLTS/DIV
2.0 NSEC/DIV
2 KHZ.

(NO 2 GENERATION)

- a) Output signal amplitude:
≥ +250 V TO 50 Ω
≥ 125 V TO 12.5 Ω (WITH AVX-M3)
- b) Pulse width:
≤ 4. NSEC
- c) Jitter, stability:
OK
- d) Prime power:

+15 ± 1.0 V
500 mA (MAX)

- e) *MODE PRF:*
 - a) *1 OUTPUT: 7 KHZ.*
 - b) *7 OUTPUTS: 1 KHZ.*

A handwritten signature or set of initials, possibly 'AJ', written in dark ink.