


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVN-S-2-P

S.N.: 1285

Date: APRIL 21 83



- a) Output signal amplitude:  
0 TO +20V
- b) Pulse width:  
200 TO 1500 PSEC
- c) Rise time:  
≤ 100 PSEC
- d) Fall time:  
≤ 135 PSEC
- e) PRF:  
50 - 260 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

+24 ± 1.0 V

400 mA (MAX)

100 NSEC/DIV

5 VOLTS/DIV

150 kHz

*[Signature]*