


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM-2-N-D*

S.N.: *1293*

Date: *MAK 17 83*



- a) Output signal amplitude:  
*0 TO -15 V*
- b) Pulse width:  
*0 TO 2 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV  
1.0 nSEC/DIV  
4.0 MHz*

*+24 ± 1.0V  
500 mA (MAX)*