


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PUM-2-N-D-PS*

S.N.: *1313*

Date: *APRIL 25 83*



a) Output signal amplitude:  
*0 TO -15V*

b) Pulse width:  
*0 TO 2 nSEC*

c) Rise time:  
*≤ 100 pSEC*

d) Fall time:  
*≤ 135 pSEC*

e) PRF:  
*0 TO 25 MHz*

f) Jitter, stability:  
*OK*

g) Prime power:

*10 VOLTS/DIV  
1.0 nSEC/DIV  
4.0 MHz*

*120/240 V,  
50-60 Hz.  
UNIT SHIPPED  
WITH VOLTAGE  
SELECTOR CARD IN  
THE 240 V POSITION*