


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AN-1*

S.N.: *1333*

Date: *May 29 83*



- a) Output signal amplitude:
25 TO +100 AMPS
- b) Pulse width:
25 NSEC
- c) Rise time:
- d) Fall time:
NA
- e) PRF:
NA
- f) Jitter, stability:
0 TO 1 KHZ
- g) Prime power:
OK

64 AMPS/DIV. +15 ± 1.0 V
10 NSEC/DIV. 140 mA (MAX.)
1 KHZ.

[Signature]