


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMN-1CP*

S.N.: *1337*

Date: *Aug 1 83*



*10 MHz
1.0 nsec/div
5 volts/div*

a) Output signal amplitude:

0 TO +5V

b) Pulse width:

0 TO 6 nsec

c) Rise time:

≤ 100 psec

d) Fall time:

≤ 200 psec

e) PRF:

0 TO 50 MHz

f) Jitter, stability:


OK

g) Prime power:

110/220V,

50-60 Hz.

[Signature]



*50 MHz.
10 nsec/div
5 volts/div.*