


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVM-2-N*

S.N.: *1349*

Date: *JUNE 27 83*



- a) Output signal amplitude:  
*0 TO 15V*
- b) Pulse width:  
*0 TO 2 nSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS / DIV*  
*1.0 nSEC / DIV*  
*4.0 MHz.*

*+ 24 ± 1.0 VOLTS*  
*500 mA (MAX)*

*AK*