


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVL-2A-N

S.N.: 1350

Date: APRIL 29 83



100 VOLTS/DIV  
10 NSEC/DIV  
4 KHZ.

- a) Output signal amplitude:  
0 TO -170 VOLTS
- b) Pulse width:  
0 TO 100 NSEC
- c) Rise time:  
 $\leq 1$  NSEC
- d) Fall time:  
 $\leq 2$  NSEC.
- e) PRF:  
0 TO 5 KHZ.
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 ± 1.0 VOLT  
275 mA (MAX)

AK