

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVM-2-P

S.N.:

1353

Date:

APRIL 25 83



a) Output signal amplitude:

0 TO +15V

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

$+24 \pm 1.0$ V

500 mA (MAX)

10 VOLTS/DIV

1.0 NSEC/DIV

4.0 MHz