


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVUM-2-1P

S.N.: 1354

Date: April 26 83



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 2 nSEC
- c) Rise time:
 ≤ 100 pSEC
- d) Fall time:
 ≤ 135 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 nSEC/DIV
4.0 MHz

+24 ± 1.0 V
500 mA (MAX)